

Influence of edge enhancement applied in endoscopic systems on sharpness and noise (Erratum)

Geert Geleijnse^a and Bernd Rieger^b

^aErasmus University Medical Center, Department of Ear, Nose, & Throat, Rotterdam, The Netherlands

^bDelft University of Technology, Department of Imaging Physics, Faculty of Applied Sciences, Delft,
The Netherlands

[DOI: [10.1117/1.JBO.28.7.079801](https://doi.org/10.1117/1.JBO.28.7.079801)]

This article [*J. Biomed. Opt.* **27**(10), 106001 (2022) doi: [10.1117/1.JBO.27.10.106001](https://doi.org/10.1117/1.JBO.27.10.106001)] was originally published on 6 October 2022 with an error in Equation 4.

Original

$$\sigma_V(Y, R, G) = \sqrt{\sigma(Y)^2 + 0.278 \cdot \sigma(Y - R)^2 + 0.088 \cdot \sigma(Y - G)^2}$$

Corrected

$$\sigma_V(Y, R, B) = \sqrt{\sigma(Y)^2 + 0.279 \cdot \sigma(Y - R)^2 + 0.088 \cdot \sigma(Y - B)^2}$$

The authors checked their software and found that they had used the correct equation there, so the reported results were unaffected by the typo. The article was corrected and republished on 8 July 2023.

